

Product Main Number 199990991501



## Examination gloves, ABENA Classic, M, clear, vinyl, powder-free

- ✓ Smooth surface
- ✓ No latex proteins or accelerators



### Product Description

Examination gloves with a smooth surface that do not contain latex proteins or accelerators. Suitable for handling non-fatty foods and short cleaning tasks without strong chemicals.



### Specifications

Base name	Examination gloves
Brand	ABENA
Sub-Brand	Classic
Size	M
Colour	Clear
Properties	Latex free, low grip, rolled cuff, smooth
Features	Powder-free
Single or multiple use	Single use
Material	Vinyl
Ingredient list	PVC, DINP, Viscosity reducer JNJ, PU-coating, Ca-Zn Stabilizer
Length/depth	240 mm
Width	95 mm
Thickness	Min. 0,07 mm
Certificates	Food contact materials. CAT III. MD. UKCA.
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	EN ISO 374-5:2016 Virus, EN ISO 374-1:2016 Type B KPT, AQL 1.5, EN 420
Directives, regulations and acts	(EU) No 10/2011, (EC) No 1935/2004, (EC) No 2023/2006, (EU) 2016/425, UK MDR 2002
Safety Instructions And Warnings	Avoid contact with fatty food products as the plasticizers in the glove can migrate to food during prolonged contact. Contain PVC and plasticizers.
Shelf Life	5 years
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

### Instructions for use/application

Examine the gloves for defects before use.

**Packaging data**

Unit	Contains	Length	Width	Height	EAN
cll	1000 pcs	32 cm	25 cm	25 cm	5713571002278
pck	100 pcs	24 cm	12 cm	6 cm	5713571002261
pcs	1 pcs				



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. The showing of the CE logo for chemical protective gloves requires that tests are carried out in accordance with test standards specified in EN ISO 374-1: 2016 + A1: 2018 - such as EN 16523-1: 2015 + A1: 2018 to determine the resistance to the permeation of chemicals. The results of this test determine the relevant pictogram symbols that can be used on the packaging and labeling.



The UKCA mark guarantees that a product is safe to use and complies with all safety precautions in line with product standards and legislation of the UK.



The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.



The product does not contain latex.



The glove has a smooth surface.



The glove is 240 mm long.



The glove has a low grip.



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.