



Product Main Number 290254



Excellent

Examination gloves, ABENA Excellent, XL, off-white, stretch vinyl, powder-free

- ✓ Smooth surface
- ✓ No latex proteins or accelerators



Product Description

Examination glove with smooth surface and contains no latex proteins or accelerators. Suitable for handling non-fatty food products and minor cleaning tasks without strong chemicals.



LOW GRIP

AQL 1.5

EN EN 420 455

Specifications

Base name	Examination gloves
Brand	ABENA
Sub-Brand	Excellent
Size	XL
Secondary color	Off-white
Properties	Latex free, low grip, rolled cuff, smooth
Features	Powder-free
Single or multiple use	Single use
Material	Stretch vinyl
Ingredient list	DOTP, Polyurethane, Polysiloxane, Silica, Water, Polyvinyl Chloride, Organic Phosphite, Organic Solvent, Antioxidant, Titanium Dioxide, Aluminum Hydroxide, Amorphous Silica, D70
Length/depth	240 mm
Width	115 mm
Thickness	Min. 0,05 mm
Certificates	CE. Food contact materials. MD.
CE Category (Personal Protective Equipment)	CAT I
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.5, EN 420, EN 455
Directives, regulations and acts	(EU) No 10/2011, (EC) No 1935/2004, (EC) No 2023/2006, (EU) 2016/425, BEK nr 681 af 25/05/2020, MDR (EU) 2017/745
Safety Instructions And Warnings	Avoid contact with fatty food products, as the plasticisers in the glove can migrate to food during prolonged contact. Contains PVC and plasticisers.
Shelf Life	3 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

Instructions for use/application

Examine the gloves for flaws and defects before use

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	1000 pcs	32 cm	25 cm	25 cm	5703538262708
pck	100 pcs	24 cm	12 cm	6 cm	5703538262715
pcs	1 pcs				



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products and medical equipment.



The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.



The product does not contain latex.



The glove has a smooth surface.



The glove is 240 mm long.



The glove has a low grip.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.