**Product Main Number** 290482

ABENA Classic

# Examination gloves, ABENA Classic Sensitive, S, black, nitrile, powder-free

- ✓ Elastic
- ✓ Extra soft
- ✓ Fits to the hand
- ✓ Thin, for good sensitivity
- ✓ Black

MD CE CAT III

## Product Description

Elastic examination glove that fits to the shape of the hand during use and provides good sensitivity. Can be used for a variety of tasks such as cleaning, care, food handling, laboratory, medication handling, cytostatics and hormone creams. Contains no latex proteins, so safe for use in the event of latex allergy.

LATEX · MEDIUM GRIP · 240 mm · AQL 1.5 · 374-1:2016/ Type B KPT · EN ISO 374-5: 2016 VIRUS · EN 455 · EN ISO 21420

## Specifications

| | |
|---|---|
| **Base name** | Examination gloves |
| **Brand** | ABENA |
| **Sub-Brand** | Classic Sensitive |
| **Size** | S |
| **Colour** | Black |
| **Properties** | Finger textured, latex free, medium grip, rolled cuff, sensitive |
| **Features** | Powder-free |
| **Single or multiple use** | Single use |
| **Material** | Nitrile |
| **Ingredient list** | Nitrile NBR, Potassium Hydroxide, Zinc oxide, ZDEC, ZDBC, Sulphur, Titanium Dioxide, black pigment, Antifoam ELP |
| **Length/depth** | 240 mm |
| **Width** | 86 mm |
| **Weight, net** | 3 g |
| **Thickness** | 0,05 mm |
| **Certificates** | CE. Food contact materials. CAT III. MD. |
| **CE Category (Personal Protective Equipment)** | CAT III |
| **CE Class (Medical Devices)** | Class I |
| **Product or test standards** | AQL 1.5, EN ISO 374-1:2016 Type B KPT, EN ISO 374-5:2016 Virus, EN 455, EN ISO 21420:2020 |
| **Directives, regulations and acts** | (EC) No 1935/2004, (EC) No 2023/2006, (EU) 2016/425, BEK nr 681 af 25/05/2020, MDR (EU) 2017/745 |
| **Shelf Life** | 5 years |
| **Storage Instructions** | Store dry, clean and at room temperature. |
| **Product Disposal Instructions** | Can be disposed of with normal household waste sorted according to local regulations. |
| **Packaging Disposal Instructions** | Can be disposed of with normal household waste sorted according to local regulations. |

29048203

## Instructions for use/application

Examine the gloves for flaws and defects before use.

## Packaging data

| Unit | Contains | Length | Width | Height | EAN |
|---|---|---|---|---|---|
| cll | 1000 pcs | 26 cm | 25.6 cm | 25.4 cm | 5703538355417 |
| pck | 100 pcs | 24.2 cm | 12.2 cm | 5 cm | 5703538355424 |
| pair | 2 pcs | | | | |
| pcs | 1 pcs | | | | |

MD

Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.

CE

The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products and medical equipment.

CAT III

Third-party type approval is required for all Category III personal protective equipment (PPE) products. The showing of the CE logo for chemical protective gloves requires that tests are carried out in accordance with test standards specified in EN ISO 374-1: 2016 + A1: 2018 - such as EN 16523-1: 2015 + A1: 2018 to determine the resistance to the permeation of chemicals. The results of this test determine the relevant pictogram symbols that can be used on the packaging and labeling.

The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.

---

LATEX

The product does not contain latex.

The glove has a finger textured surface.

MEDIUM GRIP

The glove has a medium grip.

240 mm

The glove is 240 mm long.

374-1:2016/ Type B KPT

The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.

EN ISO 374-5: 2016 VIRUS

The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.

EN 455

EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

EN ISO 21420

This standard specifies the general requirements and relevant test procedures for glove design and construction, innocuousness, comfort and efficiency, as well as the marking and information supplied by the manufacturer applicable to all protective gloves.

29048203

Gloves | Nitrile
POWDER-FREE | MEDICAL EXAMINATION GLOVES
S
6-7
Classic
SENSITIVE
100 pcs.

Gloves | Nitrile
POWDER-FREE | MEDICAL EXAMINATION GLOVES
S
6-7
Classic
SENSITIVE
100 pcs.

29048203