



Product Main Number 3958

ABENA *Excellent*


### Examination gloves, ABENA Excellent, S, blue, nitrile, powder-free

- ✓ Elastic
- ✓ Fits to hand
- ✓ No latex proteins



#### Product Description

Nitrile examination glove, strong quality. Elastic and adapts to the shape of the hand during use. Can be used for a variety of tasks such as cleaning, care, food handling, laboratory, medication handling, cytostatics and hormone creams. Contains no latex proteins and therefore safe for use in the event of latex allergy



#### Specifications

Base name	Examination gloves
Brand	ABENA
Sub-Brand	Excellent
Size	S
Colour	Blue
Properties	Extra strong, finger textured, latex free, medium grip, rolled cuff
Features	Powder-free
Single or multiple use	Single use
Material	Nitrile
Ingredient list	Nitrile NBR, Titanium Dioxide, Zinc oxide, Sulfur, ZDBC, Ammonium, Antioxidant, Calcium stearate, Calcium Nitrate, Chlorine, T 2227 blue, PL 1508 blue
Length/depth	240 mm
Width	86 mm
Thickness	0,09 +/- 0,02 mm
Certificates	CE. Food contact materials. CAT III. MD.
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.0, EN ISO 374-1:2016 Type B JKPT, EN 421:2010 Radioactive Contamination, EN 455, EN ISO 21420:2020, EN ISO 374-5:2016 Virus, EN ISO 374-4:2019
Directives, regulations and acts	(EC) No 1935/2004, (EC) No 2023/2006, (EU) 2016/425, BEK nr 681 af 25/05/2020, MDR (EU) 2017/745
Shelf Life	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

### Instructions for use/application

Examine the gloves for flaws and defects before use.

### Packaging data

Unit	Contains	Length	Width	Height	EAN
cll	1000 pcs	29.1 cm	24.9 cm	28.1 cm	5703538118357
pck	100 pcs	24 cm	13.5 cm	5.6 cm	5703538118333
pair	2 pcs				
pcs	1 pcs				



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products and medical equipment.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. The showing of the CE logo for chemical protective gloves requires that tests are carried out in accordance with test standards specified in EN ISO 374-1: 2016 + A1: 2018 - such as EN 16523-1: 2015 + A1: 2018 to determine the resistance to the permeation of chemicals. The results of this test determine the relevant pictogram symbols that can be used on the packaging and labeling.



The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.



The product does not contain latex.



The glove has a finger textured surface.



The glove has a medium grip.



The glove is 240 mm long.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals n-heptane, sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

**EN ISO 21420**

This standard specifies the general requirements and relevant test procedures for glove design and construction, innocuousness, comfort and efficiency, as well as the marking and information supplied by the manufacturer applicable to all protective gloves.



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.

